

# EU DECLARATION OF CONFORMITY<sup>1</sup>



Number<sup>2</sup>

E\_GMZ\_27HQ710S\_DOC\_202207290001

## Name and address of the Manufacturer<sup>3</sup>

LG Electronics Inc.  
168, Suchul-daero, Gumi-si,  
Gyeongsangbuk-do,  
39368 Republic of Korea

(Head Office)  
LG Electronics Inc.  
LG Twin Towers, 128 Yeoui-daero, Yeongdeungpo-gu,  
Seoul, 07336, Korea

This declaration of conformity is issued under the sole responsibility of the manufacturer<sup>4</sup>

## Object of the declaration<sup>5</sup>

### Product information<sup>6</sup>

Product Name  
*Medical Monitor*

Model Name  
*27HQ710S*

### Additional information<sup>7</sup>

*CND Code: Z110601  
Adaptor: DA-120D19  
Serial Number: YMMKCXXXXXXXX (Y:year, MM:month, KC:manufacturing site, XXXXXXXX:random)  
SRN: Manufacturer (KR-MF-000001897), Authorized Representative (NL-AR-000000546)  
Basic UDI-DI: 8806098MONITOR-SURGICAL62*

The device that is covered by the present declaration is in conformity with this regulation and any other relevant Union legislation that provides for the issuing of an EU declaration of conformity:<sup>8</sup>

- References to the relevant harmonized standards used or references to the technical specifications in relation to which conformity is declared<sup>9</sup>

Medical Device Regulation 2017/745 as amended by regulation 2020/561

*CLASSIFICATION : Class I by Rule 13 of Annex VIII, Medical Device Regulation 2017/745 as amended by regulation 2020/561  
CONFORMITY ASSESSMENT : Performed according to Annex II, Annex III, Annex IV and Annex V, Medical Device Regulation 2017/745 as amended by regulation 2020/561*

<i>EN ISO 13485:2016/A11:2021</i>	<i>EN 60601-1-2:2015</i>	<i>EN 60601-1:2006/A2:2021</i>
<i>EN ISO 14971:2019/A11:2021</i>	<i>EN 55011:2009/A1:2010</i>	<i>EN 62304:2006/A1:2015</i>
<i>ISO/TR 24971:2020</i>		<i>EN 62366-1:2015/A1:2020</i>
<i>ISO/TR 20416:2020</i>		<i>EN 60601-1-6:2010/A2:2021</i>
<i>EN ISO 15223-1:2021</i>		
<i>EN ISO 20417:2021</i>		

RoHS Directive 2011/65/EU

*DIN EN IEC 63000:2019-05*

### Intended Purpose

*This Medical Monitor is intended to provide color video displays and images from medical equipment which include laparoscope and endoscopy systems for surgery and various medical imaging systems.*

The notified body<sup>10</sup>

N/A

performed

N/A

and issued the certificate

N/A

### Additional information<sup>7</sup>

#### Factory Information:

LG Electronics Inc.  
168, Suchul-daero, Gumi-si, Gyeongsangbuk-do, 39368 Republic of Korea  
LG Electronics Nanjing New Technology Co., Ltd.  
NO.346, Yaoxin Road, Economic & Technical Development Zone, Nanjing, 210038 China

Signed for and on behalf of:<sup>11</sup> LG Electronics Inc.

Authorized Representative:

LG Electronics European Shared Service Center B.V.  
Krijgsman 1, 1186 DM Amstelveen, The Netherlands

Name and Surname / Function:

Yun Hee Yang / Director

Date of issue:

29. July. 2022

